

Essential Spares & Consumables for operation of Laser System

Table 2: Technical specifications of Essential Spares and Consumables

No.	Parameter	Specification/Quantity	Remarks to Party
1	Essential spares (items not covered under warranty) for the operation of the offered Laser System for 5000 hours	Specifications of spares suited to the offered laser system	<i>List of items and quantity should be provided. Shelf life has to be clearly stated.</i>
2	Flashlamps for Oscillator , required for the operation of the offered Laser System	Quantity: 15 Nos. Specifications of flashlamps (lamp life, shelf life, operational/storage conditions, etc.) suited to the offered laser system.	<i>Shelf life and storage conditions have to be clearly stated.</i>
3	Flashlamps for Amplifiers* , required for the operation of the offered Laser System	Quantity: 30 Nos. Specifications of flashlamps (lamp life, shelf life, operational/storage conditions, etc.) suited to the offered laser system	<i>Shelf life and storage conditions have to be clearly stated.</i> <i>*Number of amplifiers in the offered laser is to be clearly specified.</i>
4	Cartridges/Filters required for the operation of the offered Laser System	Quantity: 10 Nos. Specifications of cartridges/filters suited to the offered laser system	<i>Shelf life and storage conditions have to be clearly stated.</i>
5	Any other essential consumables required for the operation of the offered Laser System for 5000 hours	Specifications suited to the offered laser system	<i>List of items and quantity should be provided.</i>