| General Description | Shore D testing unit conforming to ASTM |
|--------------------------|--|
| · | D 2240 standard to be installed on the |
| | multi unit tester. |
| Range | 0 to 100 Sh.D |
| Resolution | 0.01 Sh.D |
| Accuracy | ± 1 Sh.D |
| Calibration of equipment | The instrument shall be supplied with |
| | calibration certificate with traceability. The |
| | certificate includes follows. |
| | Calibration of the force applied by the |
| | spring vs indentation 2. Calibration of the |
| | hardness reading vs indentation |
| | 3. Calibration of the test time. |
| Standard blocks | Samples required for verifying minimum 5 |
| | measurements with standard hardness |
| | values of 20-90 Sh.D. The samples should |
| | be supplied with 1. Calibration |
| | certificate and traceability. |
| | Single identification label |
| | 3. Wooden protection box |
| Range of application | Suitable for hardness measuring on flat |
| | and curved surfaced specimens with |
| | diameter 35 mm and thickness 6 mm |
| | prescribed under standards. |