

Annexure 1

Capacitive Vacuum Transducer Type 1

Range: 1100 hpa

Type: CMR 371

Input: 14 to 30V DC, 12 W LPS

Terms & conditions

1. The Party shall supply, install and commission the above transducer in the Vacuum calibration system
2. The Party shall furnish NABL accredited calibration certificate along with the item
3. Warranty: 1 year