

Specifications for Terahertz Non-Destructive Evaluation System

3. Specifications for spare THz focusing lens

- | | | |
|------|------------------|---|
| 3.1. | Working Distance | 50 mm |
| 3.2. | Diameter | 50 mm (typical) |
| 3.3. | Lens mount | Easily replaceable magnetically held helicoidal mount |