

After pre-bid meeting with vendors to bring out more clarity following points of RFP are updated-

Sr. no.	RFP ref no.	Specification details
1.	Revised Fig 1, 3 & 4 (pg no: 6 & 7)	Signal Generator renamed as “Signal Generator 1” in Fig 1, 3 & 4.
2.	Addendum Sr No 2.4.6 (pg no : 14)	An accredited calibration report shall be provided for all antennas.
3.	Revised Sr No 3.7.1.1.16 & 3.7.1.2.16 (pg no : 25 & 26)	The software should support test instruments like EMI Receivers, Signal Generators, Function Generators, Power Meters and RF Amplifiers of standard makes with available remote access/control using GPIB (IEEE 488.2) / LAN / USB interfaces and SCPI command sets, so that in case of any failure/change in instruments, the same can be replaced with an equivalent instrument of other make and model with GPIB/LAN/USB Interface. If software support for any of the above listed instruments is from a single make only, it shall be treated as Non Compliance.
4.	Revised Sr No 3.8 (pg no : 26 & 27)	VI. One RF switch matrix will be co shared between 10 kHz – 40GHz RE test setup and 10 kHz -1GHz RS test setup. VII. Second RF switch matrix will used in mobile microwave rack for RS test. Qty. : 2 (1 for each)
5.	Revised Sr No 14 in Annexure-2 (pg no : 30)	14. RF Switch matrix (Cater 1GHz -40GHz RS test requirement) with Qty. 1
6.	Revised Annexure-3 (pg no : 31)	Layout modified and following note added. <i>“Note: *Signal Generator 2 is SAC’s responsibility and it is not part of deliverables.”</i>
7.	Commercial terms: Only Class-I & Class-II local suppliers are eligible to participate in this tender. Further, the quotation should be in INR only. Bid either by non-local supplier or in currency other than INR will be rejected summarily.	

All the above changes are incorporated in the RFP and revised RFP also uploaded. Bidders are requested to quote accordingly. Evaluation will be done based on revised RFP and terms mentioned in above table.